

Digital Hand-held "Pocket" Refractometer

PAL-Urea







The proven choice in Europe for fast, accurate measurements of Diesel Exhaust Fluid.

PAL-Urea quickly and accurately measures the urea concentration in emissions-reducing diesel exhaust fluid additives.

Features

- Simple calibration with tap water
- Measurement in 3 seconds
- Convenient storage case
- Light & Compact, 100g
- Equipped with a strap holder option
- Extremely water resistant (IP65)
- External Light Interference (ELI)
- Automatic Temperature Compensation (ATC)
- Ergonomically designed for simple one-handed operation

Specifications

Measurement range : Urea 0.0 to 55.0% (w/w)

Resolution : Urea 0.1% Measurement accuracy : Urea ±0.3% Measurement temperature : 10 to 40°C (ATC)

Sample volume : 0.3ml Measurement time : 3 seconds

Power supply : AAA alkaline battery × 2

: Approx. 11,000 measurements Battery life

(Using new alkaline batteries)

International Protection Class : IP65 Water resistant


Dimensions and weight : 55(W) × 31(D) × 109(H)mm, 100g



Actual size

Measurement Method

1. Place the sample onto the prism surface.







3. The measurement value will be displayed in 3 seconds.



ISO9001

ASR REGISTERED



All ATAGO refractometers are designed and manufactured in Japan.



ATAGO products comply with HACCP, GMP, and GLP system standards.



Headquarters: The Front Tower Shiba Koen, 23rd Floor 2-6-3 Shiba-koen, Minato-ku, Tokyo 105-0011, Japan TEL: 81-3-3431-1943 FAX: 81-3-3431-1945 overseas@atago.net http://www.atago.net/

PATAGO U.S.A., Inc.
TEL: 1-425-637-2107 customerservice@atago-usa.com

TEL: 91-22-2833-8038 / 8076 customerservice@atago-india.com

ATAGO (THAILAND) Co., Ltd.
TEL: 662-982-8718-9 customerservice@atago-thailand.com

PATAGO BRASIL Ltda.
TEL: 55 16 3916-6000 customerservice@atago-brasil.com

ATAGO ITALIA s.r.I.
TEL: 39 2 36557267 customerservice@atago-italia.com

ATAGO CHINA Guangzhou Co., Ltd.
TEL: 86 20-38106065 info@atago-china.com

* Specifications and appearance are subject to change without notice.